

Guidelines for the Blood Transfusion Services

23.6: Component labels

<http://aws-lon-jpac.targetservers.uk/red-book/chapter-23-specification-for-the-uniform-labelling-of-blood-blood-components-and-blood-donor-samples/23-6-component-labels>

23.6: Component labels

23.6.1: General description

These labels are for use on blood collection packs and/or satellite packs. Each label will display a component description printed in bold text, a Codabar barcode and additional information. All information is printed in black on a white background. These labels may be pre-printed or produced using a demand-printed system where the information is transferred electronically from a host system.

23.6.2: Label dimensions

Label dimensions are defined below:

55 mm \pm 1 mm wide \times 55 mm +1/–3 mm deep

(range 54–56 mm wide \times 52–56 mm deep)

23.6.3: Label specification

The label must meet the following specifications:

- The barcode height must be no less than 10 mm with a 2 mm quiet zone each side of it.
- The barcode must have the eye-readable code textually displayed to accompany the barcode.
- The textual component description must be in bold characters and be exactly as is registered in the UK portfolio.
- Use of abbreviations must be authorised by the Standing Advisory Committee on Blood Components (SACBC) or JPAC groups.
- A UK JPAC agreed instruction statement must be included.
- The volume of the component must be textually displayed on the label as either the exact or as a nominal volume in millilitres (mL).
- Any storage or special instructions for storage must be included.
- The recipe for any included anticoagulant or additive must be indicated on the label in textual form.

- Where the component is part of a split component, the split should be identified as Pack 1, Pack 2 etc. textually on the label.
- If the component has a reference number (i.e. CT number), it should be included on the label in textual form with or without version number as a suffix.

Figure 23.7 shows an example of a component label layout.

23.6.4: Component barcodes

All components have an individual barcode. The barcode comprises three main elements:

- a start code 'a0'
- a five-character code to uniquely denote the component
- a stop code '3b'.

23.6.5: Component code reference table

The current component code reference table is held and managed by SACIT. The table includes:

- the text defining the component. Where possible this text is the same as that defined in Chapter 7 of these guidelines
- the start code for the component
- the code for the component, e.g. 04260
- the stop code for the component.


INSTRUCTIONS Always check patient/component compatibility/identity Inspect pack for signs of deterioration or damage Risk of adverse reaction/infection, including vCJD	
 04333	
Red Cells Leucocyte Depleted, in additive solution STORE AT 4°C ± 2°C <div style="text-align: right;">Volume XXX ml</div>	
Component collected into CPD anticoagulant. Red cells suspended in 100 millilitres of additive solution with the composition in mmol/l: Sodium chloride 150, Adenine 1.25, Glucose 45 and Mannitol 29	

Figure 23.7 Example of a component label layout

23.6.6: Allocation of new component codes

In the event of a UK Blood Establishment requiring a code for a component that will be issued on a regular basis, the following steps must be carried out:

- A request form for a new component code, including a component being trialled, must be completed and sent to SACBC. This committee will determine if the component is a 'new' component or if it is covered by an existing component registration. If it is a new component it will be reviewed by SACBC members and if agreed, the product will be accepted.
- SACBC will authorise the component to be part of the UK component portfolio and it will then be numbered with a new unused Codabar component code.